

Steam Sterilization Monitoring and System Failures: Recalls

This checklist is an excerpt from [Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices](#). It provides guidance related to situations that require recall of reprocessed medical equipment/devices and considerations to be included in recall procedures. For more information, please visit www.publichealthontario.ca or email ipac@oahpp.ca.

Inadequate reprocessing includes, but is not limited to, the following situations:

- The load contains a positive Biological Indicator (BI).²³
- An incorrect reprocessing method was used on the equipment/device.
- Steam sterilizer indicates failure to reach correct cycle parameters (e.g., time temperature, pressure).
- Chemical Indicator (CI) or monitoring tape has not changed colour.
- There is doubt about the sterility of medical equipment/devices.^{1,2}

A written procedure must be established for the recall and reprocessing of improperly reprocessed medical equipment/devices.^{1,2} All equipment/devices in each processed load must be recorded to enable tracking in the event of a recall. Facilities should consider implementing commercial instrument tracking systems to facilitate identification of patients in the event of a recall.

The recall procedures should include²:

- Designation of department and staff responsible for executing the recall.¹
- Identification of the medical equipment/devices to be recalled¹; if recall is due to a failed BI, the recall shall include the medical devices in the failed load as well as all other devices processed in the sterilizer since the last successfully sterilized load.¹
- Assessment of client/patient/resident risk.
- Procedure for subsequent notification of physicians, clients/patients/residents, other facilities and/or regulatory bodies, if indicated.
- Surveillance of clients/patients/residents, if indicated.
- Quarantine of recalled items pending the results of investigation.
- Involvement of the facility's risk manager, if applicable.

Health care settings shall have a process for receiving and disseminating medical device alerts and recalls originating from manufacturers or government agencies.¹⁴

Reusable medical devices that have been recalled due to a reprocessing failure shall be reprocessed prior to use.

Recommendations

75. If a failed chemical indicator is found, the contents of the package shall be reprocessed before use. [CSA Z314.3]
76. A procedure shall be established for the recall of improperly reprocessed medical equipment/devices. [CSA Z314.0]
77. The recall procedure shall include assessment of client/patient/resident risk and a procedure for subsequent notification of physicians, clients/patients/residents, other facilities and/or regulatory bodies if indicated. [CSA Z314.0]
78. Health care settings shall have a process for receiving and disseminating medical device alerts and recalls originating from manufacturers or government agencies. [CSA Z314.0]

References

The following references follow the order of original document. For full reference list please refer to [Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices](#).

1. CSA Group. CAN/CSA-Z314.3-09: Effective sterilization in health care facilities by the steam process. Toronto, ON: CSA Group; 2009.
2. CSA Group. CAN/CSA-Z314.0-13: Medical device reprocessing - general requirements. Toronto, ON: CSA Group; 2013.
14. CSA Group. CAN/CSA Z314.8-08: Decontamination of reusable medical devices. Toronto, ON: CSA Group; 2008.
23. Health Canada. Infection control guidelines: hand washing, cleaning, disinfection and sterilization in health care [Internet]. Ottawa, ON: Health Canada; 1998 [cited 2017 Oct 23]. Available from: http://publications.gc.ca/collections/collection_2016/aspc-phac/HP3-1-24-S8-eng.pdf